



# Certificate of Calibration

Iraq- Basra- Jaza' er, Field : North Rumela

•Phone : + 964 7833334444 • www.aberdeents.com • E-mail: akader@aberdeents.com

Date of Issue: 31 July 2023

Page 1 of 1

REQUEST NUMBER : By Contract	APPROVED BY LAB IN CHARGE QC
JOB NUMBER : QC/JN/23/00266	
CERTIFICATE NUMBER : QC230730-049	



### CUSTOMER DETAILS

Name : DPAB Co.Ltd Iraq Branch Block-9  
 Address : Basra - Iraq, Block 9  
 Rig : DQ43

### EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : Pressure Relief Valve  
 Manufacturer : N/A  
 Model : N/A  
 Serial Number : CAL-PRV-001  
 Type : Analog  
 Location : SCBA Refill Compressor  
 Range : 350 bar  
 Calibrated By : Mahdi Haleem  
 Calibration Date : 31 July 2023  
**Calibration Due : 30 July 2024**



### ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 25 ± 1°C Humidity : 44 ± 5 % RH Atm. Pressure 1011 ± 10 hPa

### CALIBRATION METHOD

Standard Method API 527 was used to test the unit under test. Test pressure was applied on inlet of unit under test and test was repeated 3 times. Identical repeatings were carried out as recommended in used standard.

### REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	CERTIFICATE NO.	SOURCE OF TRACEABILITY
Digital Pressure Gauge	1505119397	LT-221215-03	Liftek Iraq

### CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Tolerance Limit
bar	bar	bar	± bar
350	343.2	6.8	10.5

Results :

The above Value Indicates That the Instrument complies with the Specified Maximum Permissible error tolerance.

At The Time of Calibration. and ready for use.

DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

The results relate to measurements taken at nominal value only.  
 Adjustments were made prior to calibration.  
 Accessories were fitted.

~ END OF DOCUMENT ~

