



ABERDEEN COMPANY FOR TECHNICAL SERVICES

Email: workshop@aberdeenTS.com

Contact: +964782222633



CERTIFICATE OF CALIBRATION

Date of Issue: November 3, 2025

CUSTOMER DETAILS			
Name	Quality Control Iraq	Request Number	By Mail
Address	North Rumaila ,Basra, Iraq	Job Number	ATS/25/350
Department	Calibration Unit	Certificate Number	ATS-251103-01


Details of Item / Unit under calibration			
Description	Pressure Gauge	Calibration Date	November 3, 2025
Serial Number	468802	Calibration Due	November 2, 2026
Manufacturer	AMETEK	Range	0 - 5000 psi
Model	XP2i	Readability	50 psi
Type	Digital	Reference Standard	API 527

Ambient Conditions			
Ambient Temperature	27 ± 1°C	Humidity	41 ± 1 % RH

Reference Equipment Used			
Description	SERIAL NO.	Make	SOURCE OF TRACEABILITY
Pressure Gauge	54179	JAPAN	DELTA ENGINEERING

CALIBRATION TEST RESULTS			
Unit Under Calibration	Reference Instrument Reading	Deviation	Expanded Uncertainty ± (IN)
psi	psi		
1000	1002	0.002	0.001
3000	3004	0.004	0.001
5000	5007	0.007	0.001

Remarks	Calibrated & tested as per Manufacturer's instructions. And Found okay
---------	--

Calibrated By	Shaker Ahmad
Signature :	

Authorized By	Ahmad Imam
Signature:	

